


# Technical Appendix 10.3: Calibration certificates of noise instruments

		<p>MTS Calibration Ltd, The Grange Business Centre, Belasis Avenue, Billingham TS23 1LG, England Telephone: 01624 876 410</p>																																																																																																					
<h2>CERTIFICATE OF CALIBRATION</h2>		<p>Page 1 of 11 pages</p> <p>Approved Signatory:</p> <p style="text-align: center;"><i>PA Sherris</i></p> <p style="text-align: right;">Tony Sherris</p>																																																																																																					
<p>Issued by: <b>MTS Calibration Ltd</b></p>		<p>Date of Issue: <b>25 January 2019</b>      Certificate Number: <b>32812</b></p>																																																																																																					
<h3>Sound Level Meter</h3>																																																																																																							
<h3>Sound Level Meter Periodic Tests to EN 61672-3: 2013 Class 1</h3>																																																																																																							
<p><b>Client:</b> Environmental Measurements on behalf of Brendan O'Reilly Unit 12, Tallaght Business Centre Whitestown Business Park Co.Dublin 24, Ireland</p>		<p><b>Instrument Make:</b> Larson Davis <b>Instrument Model:</b> LX1L <b>Serial Number:</b> 0004570</p>																																																																																																					
<p><b>8</b></p>	<table border="0"> <thead> <tr> <th style="text-align: left;">Associated Equipment</th> <th style="text-align: left;">Make</th> <th style="text-align: left;">Model</th> <th style="text-align: left;">Serial number</th> </tr> </thead> <tbody> <tr> <td>Preamplifier</td> <td>PCB</td> <td>PRMLxT1L</td> <td>036058</td> </tr> <tr> <td>Microphone</td> <td>PCB</td> <td>377B02</td> <td>152974</td> </tr> <tr> <td>Calibrator</td> <td>Larson Davis</td> <td>CAL200</td> <td>9175</td> </tr> <tr> <td colspan="2">Calibrator supplied by</td> <td colspan="2">by MTS for this calibration</td> </tr> </tbody> </table>			Associated Equipment	Make	Model	Serial number	Preamplifier	PCB	PRMLxT1L	036058	Microphone	PCB	377B02	152974	Calibrator	Larson Davis	CAL200	9175	Calibrator supplied by		by MTS for this calibration																																																																																	
Associated Equipment	Make	Model	Serial number																																																																																																				
Preamplifier	PCB	PRMLxT1L	036058																																																																																																				
Microphone	PCB	377B02	152974																																																																																																				
Calibrator	Larson Davis	CAL200	9175																																																																																																				
Calibrator supplied by		by MTS for this calibration																																																																																																					
<p><b>Test results summary, detailed results are shown on subsequent pages.</b></p>																																																																																																							
<p>Periodic tests were performed in accordance with procedures from IEC 61672-3:2013 Class 1</p>																																																																																																							
<table border="0"> <thead> <tr> <th style="text-align: left;">Tests performed</th> <th style="text-align: left;">Section</th> <th style="text-align: left;">Results of test</th> <th style="text-align: left;">Page</th> <th style="text-align: left;">Comments</th> </tr> </thead> <tbody> <tr> <td>Calibration Certificate</td> <td>22</td> <td></td> <td>1</td> <td></td> </tr> <tr> <td>Additional information</td> <td></td> <td></td> <td>2</td> <td></td> </tr> <tr> <td>Indication with Calibrator Supplied</td> <td>10</td> <td>No Limit</td> <td>3</td> <td></td> </tr> <tr> <td>Self-Generated Noise</td> <td>11</td> <td>No Limit</td> <td>3</td> <td></td> </tr> <tr> <td>Frequency and Time-weightings at 1kHz</td> <td>14</td> <td>Complies</td> <td>3</td> <td></td> </tr> <tr> <td>Long term stability</td> <td>15</td> <td>Complies</td> <td>3</td> <td></td> </tr> <tr> <td>High stability</td> <td>21</td> <td>Complies</td> <td>3</td> <td></td> </tr> <tr> <td>Acoustic Tests</td> <td>12</td> <td>Complies</td> <td>4</td> <td></td> </tr> <tr> <td>Frequency Weighting A</td> <td>13</td> <td>Complies</td> <td>5</td> <td></td> </tr> <tr> <td>Frequency Weighting C</td> <td>13</td> <td>Complies</td> <td>6</td> <td></td> </tr> <tr> <td>Frequency Weighting Z</td> <td>13</td> <td>Complies</td> <td>7</td> <td></td> </tr> <tr> <td>Level Linearity</td> <td>16</td> <td>Complies</td> <td>8</td> <td></td> </tr> <tr> <td>Level Linearity Range Control</td> <td>17</td> <td></td> <td>n/a</td> <td>Only one range</td> </tr> <tr> <td>Tone-burst Response</td> <td>18</td> <td>Complies</td> <td>9</td> <td></td> </tr> <tr> <td>Peak C sound level</td> <td>19</td> <td>Complies</td> <td>10</td> <td></td> </tr> <tr> <td>Overload indication</td> <td>20</td> <td>Complies</td> <td>11</td> <td></td> </tr> <tr> <td colspan="5"><b>Additional tests performed</b></td> </tr> <tr> <td>Microphone</td> <td></td> <td>32814</td> <td></td> <td>See additional certificate</td> </tr> <tr> <td>Filter, third octave or octave</td> <td></td> <td>32812F</td> <td></td> <td>See additional certificate</td> </tr> </tbody> </table>				Tests performed	Section	Results of test	Page	Comments	Calibration Certificate	22		1		Additional information			2		Indication with Calibrator Supplied	10	No Limit	3		Self-Generated Noise	11	No Limit	3		Frequency and Time-weightings at 1kHz	14	Complies	3		Long term stability	15	Complies	3		High stability	21	Complies	3		Acoustic Tests	12	Complies	4		Frequency Weighting A	13	Complies	5		Frequency Weighting C	13	Complies	6		Frequency Weighting Z	13	Complies	7		Level Linearity	16	Complies	8		Level Linearity Range Control	17		n/a	Only one range	Tone-burst Response	18	Complies	9		Peak C sound level	19	Complies	10		Overload indication	20	Complies	11		<b>Additional tests performed</b>					Microphone		32814		See additional certificate	Filter, third octave or octave		32812F		See additional certificate
Tests performed	Section	Results of test	Page	Comments																																																																																																			
Calibration Certificate	22		1																																																																																																				
Additional information			2																																																																																																				
Indication with Calibrator Supplied	10	No Limit	3																																																																																																				
Self-Generated Noise	11	No Limit	3																																																																																																				
Frequency and Time-weightings at 1kHz	14	Complies	3																																																																																																				
Long term stability	15	Complies	3																																																																																																				
High stability	21	Complies	3																																																																																																				
Acoustic Tests	12	Complies	4																																																																																																				
Frequency Weighting A	13	Complies	5																																																																																																				
Frequency Weighting C	13	Complies	6																																																																																																				
Frequency Weighting Z	13	Complies	7																																																																																																				
Level Linearity	16	Complies	8																																																																																																				
Level Linearity Range Control	17		n/a	Only one range																																																																																																			
Tone-burst Response	18	Complies	9																																																																																																				
Peak C sound level	19	Complies	10																																																																																																				
Overload indication	20	Complies	11																																																																																																				
<b>Additional tests performed</b>																																																																																																							
Microphone		32814		See additional certificate																																																																																																			
Filter, third octave or octave		32812F		See additional certificate																																																																																																			
<p>The instrument was within the above specification as received - no modifications were made</p>																																																																																																							
<p>The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3: 2013 for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2: 2013, to demonstrate that the model of sound level meter fully conformed to the Class 1 specifications in IEC 61672-1: 2013, the sound level meter submitted for testing conforms to the Class 1 specifications of IEC 61672-1: 2013</p>																																																																																																							
<p>This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.</p>																																																																																																							



MTS Calibration Ltd,  
The Grange Business Centre,  
Belasis Avenue,  
Billingham TS23 1LG,  
England  
Telephone: 01624 876 410

## CERTIFICATE OF CALIBRATION

Page 1 of 11 pages

Issued by: **MTS Calibration Ltd**

Approved Signatory:

Date of Issue: **24 January 2019** Certificate Number: **32818**

**Tony Sherris**

### Sound Level Meter

### Sound Level Meter Periodic Tests to EN 61672-3: 2013 Class 1

**Client:** Environmental Measurements on behalf of Brendan O'Reilly  
Unit 12, Tallaght Business Centre  
Whitestown Business Park  
Co.Dublin 24, Ireland

**Instrument Make:** Larson Davis  
**Instrument Model:** LxT1L  
**Serial Number:** 0004643

**10**

Associated Equipment	Make	Model	Serial number
Preamplifier	PCB	PRMLxT1L	042742
Microphone	PCB	377B02	173111
Calibrator	Larson Davis	CAL200	9175
Calibrator supplied by	by MTS for this calibration		

### Test results summary, detailed results are shown on subsequent pages.

Periodic tests were performed in accordance with procedures from IEC 61672-3:2013 Class 1

Tests performed	Section	Results of test	Page	Comments
Calibration Certificate	22		1	
Additional information			2	
Indication with Calibrator Supplied	10	No Limit	3	
Self-Generated Noise	11	No Limit	3	
Frequency and Time-weightings at 1kHz	14	Complies	3	
Long term stability	15	Complies	3	
High stability	21	Complies	3	
Acoustic Tests	12	Complies	4	
Frequency Weighting A	13	Complies	5	
Frequency Weighting C	13	Complies	6	
Frequency Weighting Z	13	Complies	7	
Level Linearity	16	Complies	8	
Level Linearity Range Control	17		n/a	Only one range
Tone-burst Response	18	Complies	9	
Peak C sound level	19	Complies	10	
Overload indication	20	Complies	11	
<b>Additional tests performed</b>				
Microphone		32820		See additional certificate
Filter, third octave or octave		32818F		See additional certificate

The instrument was within the above specification as received - no modifications were made

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3: 2013 for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2: 2013, to demonstrate that the model of sound level meter fully conformed to the Class 1 specifications in IEC 61672-1: 2013, the sound level meter submitted for testing conforms to the Class 1 specifications of IEC 61672-1: 2013

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.



MTS Calibration Ltd,  
The Grange Business Centre,  
Belasis Avenue,  
Billingham TS23 1LG,  
England  
Telephone: 01624 876 410

# CERTIFICATE OF CALIBRATION

Page 1 of 11 pages

Issued by: **MTS Calibration Ltd**

Approved Signatory:

Date of Issue: **25 January 2019** Certificate Number: **32815**

**Tony Sherris**

## Sound Level Meter

### Sound Level Meter Periodic Tests to EN 61672-3: 2013 Class 1

**Client:** Environmental Measurements on behalf of Brendan O'Reilly  
Unit 12, Tallaght Business Centre  
Whitestown Business Park  
Co.Dublin 24, Ireland

**Instrument Make:** Larson Davis  
**Instrument Model:** LxT1L  
**Serial Number:** 0004647

**9**

Associated Equipment	Make	Model	Serial number
Preamplifier	PCB	PRMLxT1L	042725
Microphone	PCB	377B02	171552
Calibrator	Larson Davis	CAL200	9175
Calibrator supplied by	by MTS for this calibration		

### Test results summary, detailed results are shown on subsequent pages.





Periodic tests were performed in accordance with procedures from IEC 61672-3:2013 Class 1

Tests performed	Section	Results of test	Page	Comments
Calibration Certificate	22		1	
Additional information			2	
Indication with Calibrator Supplied	10	No Limit	3	
Self-Generated Noise	11	No Limit	3	
Frequency and Time-weightings at 1kHz	14	Complies	3	
Long term stability	15	Complies	3	
High stability	21	Complies	3	
Acoustic Tests	12	Complies	4	
Frequency Weighting A	13	Complies	5	
Frequency Weighting C	13	Complies	6	
Frequency Weighting Z	13	Complies	7	
Level Linearity	16	Complies	8	
Level Linearity Range Control	17	n/a		Only one range
Tone-burst Response	18	Complies	9	
Peak C sound level	19	Complies	10	
Overload indication	20	Complies	11	
<b>Additional tests performed</b>				
Microphone		32817		See additional certificate
Filter, third octave or octave		32815F		See additional certificate

The instrument was within the above specification as received - no modifications were made

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3: 2013 for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2: 2013, to demonstrate that the model of sound level meter fully conformed to the Class 1 specifications in IEC 61672-1: 2013, the sound level meter submitted for testing conforms to the Class 1 specifications of IEC 61672-1: 2013

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

	<p>MTS Calibration Ltd, The Grange Business Centre, Belasis Avenue, Billingham TS23 1LG, England Telephone: 01624 876 410</p>																																
			0607																														
<b>CERTIFICATE OF CALIBRATION</b>		Page 1 of 1																															
Issued by: <b>MTS Calibration Ltd</b>		Approved Signatory: 																															
Date of Issue: 26 February 2019      Certificate Number: 33014U		Tony Sherris																															
<b>Sound Calibrator</b>																																	
Client: Environmental Measurements Unit 12, Tallaght Business Centre Whitestown Business Park Co.Dublin 24, Ireland																																	
The measured values were:																																	
<b>Larson Davis</b>		<b>Model CAL200</b>																															
		<b>Serial Number 5838</b>																															
<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="padding: 5px;">Output Level 1:</td> <td style="padding: 5px;">94.09 dB re 20<math>\mu</math>Pa</td> <td style="padding: 5px;"><math>\pm</math> 0.14 dB (k= 2)</td> </tr> <tr> <td style="padding: 5px;">Fundamental Frequency 1:</td> <td style="padding: 5px;">1000.50 Hz</td> <td style="padding: 5px;"><math>\pm</math> 0.11 Hz (k= 2)</td> </tr> <tr> <td style="padding: 5px;">Total Harmonic Distortion 1:</td> <td style="padding: 5px;">0.39 %</td> <td style="padding: 5px;"><math>\pm</math> 0.01 % (k= 2)</td> </tr> <tr> <td style="padding: 5px;">Output Level 2:</td> <td style="padding: 5px;">114.16 dB re 20<math>\mu</math>Pa</td> <td style="padding: 5px;"><math>\pm</math> 0.14 dB (k= 2)</td> </tr> <tr> <td style="padding: 5px;">Fundamental Frequency 2:</td> <td style="padding: 5px;">1000.48 Hz</td> <td style="padding: 5px;"><math>\pm</math> 0.11 Hz (k= 2)</td> </tr> <tr> <td style="padding: 5px;">Total Harmonic Distortion 2:</td> <td style="padding: 5px;">0.40 %</td> <td style="padding: 5px;"><math>\pm</math> 0.01 % (k= 2)</td> </tr> </table>				Output Level 1:	94.09 dB re 20 $\mu$ Pa	$\pm$ 0.14 dB (k= 2)	Fundamental Frequency 1:	1000.50 Hz	$\pm$ 0.11 Hz (k= 2)	Total Harmonic Distortion 1:	0.39 %	$\pm$ 0.01 % (k= 2)	Output Level 2:	114.16 dB re 20 $\mu$ Pa	$\pm$ 0.14 dB (k= 2)	Fundamental Frequency 2:	1000.48 Hz	$\pm$ 0.11 Hz (k= 2)	Total Harmonic Distortion 2:	0.40 %	$\pm$ 0.01 % (k= 2)												
Output Level 1:	94.09 dB re 20 $\mu$ Pa	$\pm$ 0.14 dB (k= 2)																															
Fundamental Frequency 1:	1000.50 Hz	$\pm$ 0.11 Hz (k= 2)																															
Total Harmonic Distortion 1:	0.39 %	$\pm$ 0.01 % (k= 2)																															
Output Level 2:	114.16 dB re 20 $\mu$ Pa	$\pm$ 0.14 dB (k= 2)																															
Fundamental Frequency 2:	1000.48 Hz	$\pm$ 0.11 Hz (k= 2)																															
Total Harmonic Distortion 2:	0.40 %	$\pm$ 0.01 % (k= 2)																															
<p>This measurement is valid only for the above device configured for calibration of a WS-2 microphone under the above environmental conditions. For deviation of prevailing conditions, the manufacturer's literature for the calibrator should be referred to.</p>																																	
Date of Measurements: 26 February 2019		Date of Receipt: 25 February 2019																															
<p><b>Method of calibration</b></p> <p>A Reference Calibrator was used to establish the sensitivity of the measurement chain. The same measurement chain is then used to determine the output level of the Object Calibrator by the difference between its output and that of the nominated Reference Calibrator. Four independent measurements of the third-octave band sound pressure levels produced by the Reference Calibrators and the Object Calibrator are averaged to minimise uncertainties of the calibration. The measurement chain consists of a calibrated, Reference Microphone, Reference Preamplifier and Reference Analyser.</p> <p>As well as providing a traceable measurement of the sound pressure level in the cavity of the Object Calibrator, the Calibrator's frequency and total harmonic distortion are also measured. Frequency is determined from the average of four independent measurements using a multimeter. The total harmonic distortion is measured from the average of three independent measurements by third octave analysis, subtracting the level of the fundamental frequency from the sum of the combined harmonics in the frequency band to 20kHz. The complete procedure is detailed in the MTS Calibration Ltd work procedure WP01.</p> <p>The sound pressure level generated by the calibrator in its WS2 configuration was measured by reference to B&amp;K Model 4133 Microphone and reference Sound Calibrator as shown in the Test Equipment section below.</p> <p>The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k (individually calculated as above), providing a coverage probability of approximately 95%. The uncertainty evaluation has been calculated in accordance with the current version of UKAS publication M3003. The uncertainty quoted for the Distortion Measurement is the Distortion Percentage as measured, multiplied by our Uncertainty as calculated for the individual measurement or our CMC, whichever is the larger.</p>																																	
<p><b>Measurement Conditions:</b></p> <table style="width: 100%; border-collapse: collapse;"> <tr> <td style="text-align: right;">Temperature</td> <td style="text-align: center;">22</td> <td style="text-align: center;">°C</td> <td style="text-align: center;"><math>\pm</math> 1 °C</td> </tr> <tr> <td style="text-align: right;">Atmospheric Pressure</td> <td style="text-align: center;">1032</td> <td style="text-align: center;">mBar</td> <td style="text-align: center;"><math>\pm</math> 2 mBar</td> </tr> <tr> <td style="text-align: right;">Relative Humidity</td> <td style="text-align: center;">42</td> <td style="text-align: center;">%</td> <td style="text-align: center;"><math>\pm</math> 5 %</td> </tr> </table>				Temperature	22	°C	$\pm$ 1 °C	Atmospheric Pressure	1032	mBar	$\pm$ 2 mBar	Relative Humidity	42	%	$\pm$ 5 %																		
Temperature	22	°C	$\pm$ 1 °C																														
Atmospheric Pressure	1032	mBar	$\pm$ 2 mBar																														
Relative Humidity	42	%	$\pm$ 5 %																														
<p><b>Test Equipment used during this calibration:</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="text-align: left;">Equipment</th> <th style="text-align: left;">Manufacturer</th> <th style="text-align: left;">Model</th> <th style="text-align: left;">Serial No.</th> <th style="text-align: left;">Traceability Ref.</th> <th style="text-align: left;">Calibration Due</th> </tr> </thead> <tbody> <tr> <td>Reference Calibrator</td> <td>Brüel &amp; Kjær</td> <td>4231</td> <td>3014620</td> <td>TE 133</td> <td>Oct-21</td> </tr> <tr> <td>Multimeter</td> <td>HP</td> <td>34401A</td> <td>36146A63804</td> <td>TE 105</td> <td>Jul-19</td> </tr> <tr> <td>Signal Generator (set 1)</td> <td>HP</td> <td>33120A</td> <td>US36016577</td> <td>TE 111</td> <td>Jul-19</td> </tr> <tr> <td>Real-Time Analyser (set 1)</td> <td>Larson Davis</td> <td>2900</td> <td>0492</td> <td>TE 108</td> <td>Jul-19</td> </tr> </tbody> </table>				Equipment	Manufacturer	Model	Serial No.	Traceability Ref.	Calibration Due	Reference Calibrator	Brüel & Kjær	4231	3014620	TE 133	Oct-21	Multimeter	HP	34401A	36146A63804	TE 105	Jul-19	Signal Generator (set 1)	HP	33120A	US36016577	TE 111	Jul-19	Real-Time Analyser (set 1)	Larson Davis	2900	0492	TE 108	Jul-19
Equipment	Manufacturer	Model	Serial No.	Traceability Ref.	Calibration Due																												
Reference Calibrator	Brüel & Kjær	4231	3014620	TE 133	Oct-21																												
Multimeter	HP	34401A	36146A63804	TE 105	Jul-19																												
Signal Generator (set 1)	HP	33120A	US36016577	TE 111	Jul-19																												
Real-Time Analyser (set 1)	Larson Davis	2900	0492	TE 108	Jul-19																												
<p>This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.</p>																																	